

Traceable Certificate

201 Wolf Drive • P.O. Box 87 • Thorofare, NJ 08086-0087 • Phone:856-686-1600 • Fax:856-686-1601 • www.troemner.com • e-mail: troemner@troemner.com

End User

Henry Troemner LLC
201 Wolf Drive
Thorofare, NJ 08086

Page 2 of 2

Serial Number: 0101888
Certificate Number: 123456
Date of Calibration: 3-Jun-05

No As Left Data: Intital Calibration

DC Voltage in Volts

Range	Nominal	As Found	Min	Max
0.320	0.300	0.299	0.299	0.301
3.2	3.0	2.997	2.987	3.013
32	10	9.99	9.950	10.050
32	-10	-9.99	-10.050	-9.950
32	20	19.98	19.910	20.090
32	-20	-19.98	-20.090	-19.910
32	30	29.98	29.870	30.130
32	-30	-29.98	-30.130	-29.870
320	300	299.7	298.4	301.3
1000	950	949	943	957

AC Voltage in Volts

Range	Nominal	As Found	Min	Max
3.2	3.0@50Hz	2.997	2.938	3.062
	3.0@400Hz	2.991	2.938	3.062
32	30@50Hz	29.98	29.38	30.62
	30@499Hz	30.00	29.38	30.62
	30@1kHz	29.97	29.38	30.62
320	300@50Hz	299.7	293.8	306.2
	300@1kHz	299.7	293.8	306.2
750	700@50Hz	699	684	716
	700@400Hz	699	684	716
	700@1kHz	699	684	716

Resistance in Kohms

Range	Nominal	As Found	Min	Max
0.320	0.100	0.0999	0.0993	0.1007
3.2	1.0	1.000	0.994	1.006
32	10	10.00	9.94	10.06
320	100	100.0	99.4	100.6
3200	1000	1000	994	1006
32000	10000	10010	9790	10210

DC Current in Amps

Range	Nominal	As Found	Min	Max
10	9	8.96	8.84	9.16

AC Current in Amps

Range	Nominal	As Found	Min	Max
10	9.0@50Hz	8.98	8.75	9.25
10	9.0@1kHz	9.00	8.75	9.25