


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 <p><b>0516</b></p> <p>Accredited to <b>ISO/IEC 17025:2005</b></p>	<h3>Henry Troemner, LLC</h3> <p><b>Issue No: 016    Issue date: 18 August 2010</b></p>	
	<p><b>201 Wolf Drive</b> <b>PO Box 87</b> <b>Thorofare</b> <b>NJ 08086</b> <b>USA</b></p>	<p><b>Contact: Mr J Moran</b> <b>Tel: +00 1 856-686-1600 x261</b> <b>Fax: +00 1 856-686-1601</b> <b>E-Mail: troemner@troemner.com</b> <b>Website: www.troemner.com</b></p>
<p><b>Calibration performed at the above address only</b></p>		

### DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( $k=2$ )	Remarks
MASS	(g)	(mg)	For nominal values of 10 kg and below the listed Calibration and Measurement Capability is for measurements made on standards which are constructed in accordance with the principles contained in OIML Recommendation 111 for weights of class E1.
	50 000	5.0	
	30 000	3.0	
	25 000	3.0	
	20 000	2.0	
	10 000	0.50	
	5 000	0.25	
	3 000	0.15	
	2 000	0.10	
	1 000	0.045	
	500	0.025	
	300	0.020	
	200	0.015	
	100	0.007 0	
	50	0.004 0	
	30	0.003 5	
	20	0.003 0	
	10	0.002 0	
	5	0.001 5	
	3	0.000 85	
	2	0.000 85	
	1	0.000 45	
	0.5	0.000 45	
	0.3	0.000 45	
	0.2	0.000 45	
	0.1	0.000 45	
	0.05	0.000 45	
	0.03	0.000 45	
	0.02	0.000 45	
	0.01	0.000 45	
	0.005	0.000 35	
	0.003	0.000 35	
	0.002	0.000 35	
	0.001	0.000 30	
END			