

ISO/IEC 17025 Calibration Certificate

201 Wolf Drive • P.O. Box 87 • Thorofare, NJ 08086-0087 • Phone: 856-686-1600 • Fax: 856-686-1601 • www.troemner.com • e-mail: troemner@troemner.com

SECTION 1: NAME AND ADDRESS OF CUSTOMER

Page 1 of 2

End User
Troemner
201 Wolf Drive
Thorofare N.J. 08086

Client
Troemner
201 Wolf Drive
Thorofare N.J. 08086

SECTION 2: APPROVED SIGNATORY

SECTION 3: PERSON PERFORMING WORK

Lynn Dickerson : Metrologist 

Malcolm Jones

SECTION 4: CERTIFICATE INFORMATION

Description of DUT	: Pressure Indicator & Sensor	Pressure Comparison
Order Number	: 123456	Serial Number : 41234 & PPM-1234
Manufacturer	: Heise & Ashcroft	Certificate Number : 12345-LowPRESS
Model	: PM & PPM-1	Date of Calibration : May 03 2017
Range	: -12.500 to 12.500 inH ₂ O	Date Received : Apr 28 2017
Measurement Mode	: Differential	Date Issued : May 03 2017

SECTION 5: ENVIRONMENTAL CONDITIONS DURING TEST

Temperature: 21.54 C Ambient Pressure: 759.4 mmHg Relative Humidity: 40.2 %RH

SECTION 6: PERTINENT INFORMATION

STANDARDS

Manufacturer	Model	Serial Number	Range	Cal. Due Date
Ruska	7250LP	68263	0.0000 to 30.0000 inH ₂ O	Nov 01 2017

The gauge calibrated for this report has been calibrated in accordance with Troemner's calibration procedure. PCP-CAL.

This calibration also meets specifications as outlined in ISO/IEC 17025, ANSI/NCSL Z540-1-1994, and applicable documents.

This comparison calibration was performed in Troemner's Pressure Certification Laboratory at 201 Wolf Drive, Thorofare, NJ 08086. This unit was compared to laboratory standards, which are traceable to the International System of Units (SI) through Fluke Calibration Labs.

ISO/IEC 17025 Calibration Certificate

201 Wolf Drive • P.O. Box 87 • Thorofare, NJ 08086-0087 • Phone: 856-686-1600 • Fax: 856-686-1601 • www.troemner.com • e-mail: troemner@troemner.com

NAME AND ADDRESS OF CUSTOMER

Page 2 of 2

End User

Troemner
201 Wolf Drive
Thorofare N.J. 08086

Pressure Calibration
Serial Number : 41234 & PPM-1234
Certificate Number : 12345-LowPRESS
Date of Calibration : May 03 2017

SECTION 7: CALIBRATION DATA

DUT Tolerance (0.06 %Span)

As Found / As Left Data:

Test Point	Reference Pressure inH ₂ O 20C	DUT Pressure inH ₂ O 20C	Error inH ₂ O 20C	DUT Tolerance inH ₂ O 20C	Status
1	-12.4999	-12.4977	0.0022	0.0145	Pass
2	-7.4999	-7.4989	0.0010	0.0145	Pass
3	-2.5002	-2.4991	0.0011	0.0145	Pass
4	2.4997	2.4985	-0.0012	0.0145	Pass
5	7.4997	7.4999	0.0002	0.0145	Pass
6	12.4997	12.5030	0.0033	0.0145	Pass
7	7.4997	7.4981	-0.0016	0.0145	Pass
8	2.4999	2.4996	-0.0003	0.0145	Pass
9	-2.5001	-2.4989	0.0012	0.0145	Pass
10	-7.4995	-7.4990	0.0005	0.0145	Pass
11	-12.5000	-12.4989	0.0011	0.0145	Pass

Uncertainty of the Measurement : 0.0028inH₂O

The uncertainty is an error in pressure due to the measurement process, and is calculated per NIST Technical Note 1297 using a coverage factor of k = 2 (k = 2 defines an interval having a level of confidence of approximately 95 percent).

SECTION 8: ADDENDUM

Instrument ID: L1234

DUT was zeroed prior to collecting data.

Unit Conversion: 1 inH₂O@20C = 0.24864 kPa

Next Calibration Due: Nov 03 2017